
Department of Forensic Science Breath Alcohol Section Instrument Location History 1/1/1990 To 2/26/2014



Instrument Serial Number: 010453		As of 26-Feb-14	
Start Date	End Date	Agency Name	
20-Feb-14		DFS Central Lab	
06-Aug-09	20-Feb-14	Clifton Forge PD	
22-Apr-09	06-Aug-09	DFS Central Lab	
06-Nov-08	22-Apr-09	Chase City PD	
19-May-08	06-Nov-08	DFS Central Lab	

INTOX EC/IR II

Quality Assurance Worksheet

Instrument Serial Number

010453

Worksheet Start Date

2/20/2014

Location

Clifton Forge Police Department

Address

547 Main St., Clifton Forge, VA 24422

DFS Technician

Heather Stanton

License No.

18910

Laboratory

✓ On-Site

Site Specification: No detrimental environmental conditions exist.

V

Instrument Barometer (mm HG)

736

Reference Barometer (mm HG)

738

Reference Barometer(RB)Serial #

009113

RB Calibration Due

5/8/2014

Measurement Assurance Check

Standard	(sea	ievei)
		0.300

PA Target 0.290 minimum 0.281 maximum 0.298 Sample 1 Sample 2

0.296 0.296

Precision 0

sample min sample max 0.296 0.296

Sample 3 0.296

Standard (sea level) 0.100 **PA Target** 0.097 minimum 0.094 maximum 0.100 Sample 1 Sample 2 0.099

Precision 0

0.099

sample min sample max 0.099

0.099 Sample 3 0.099

Dry gas standard Lot No. (with tank no.)

AG219101-11

Replaced dry gas standard (+O-ring)

Supplies Mouthpieces

Installed at Location

Certificates of Analysis

Removed to DFS-Central

Operator Worksheet Other:

Notes:

Attempted to move instrument to new location, but analog phone line was not ready. this will

Instrument Serial Number	010453		Certification Date	
☐ Calibrated	Certified	✓ Meas	urement Assurance C	heck
✓ Instrument Test (Successful completion den	otes satisfactory condition		leshooting/Maintenance	Worksheet Completed
Standard (sea Precision 0	0.250	minimum 0.000 sample min 0.000	maximum 0.000 sample max 0.000	Sample 1 Sample 2 Sample 3
Standard (sea Precision 0	0.150	minimum 0.000 sample min 0.000	maximum 0.000 sample max 0.000	Sample 1 Sample 2 Sample 3
Standard (sea Precision 0	0.080	minimum -0.003 sample min 0.000	maximum 0.003 sample max 0.000	Sample 1 Sample 2 Sample 3
Standard (sea Precision 0	0.020	minimum -0.003 sample min 0.000	maximum 0.003 sample max 0.000	Sample 1 Sample 2 Sample 3
All measurements are in g/2 Estimation of Uncertainty of		eability records a	are located within the B	reath Alcohol Section
DFS Technician	Heath	u Start	Date	2/24/14 2/26/14 HNG
Issuing Analyst	Sed		Date	2/27/14
	min (

Intox EC/IR II QA Worksheet Issued by Calibration and Training Program Manager Issue Date: 9-December-2013 DFS Document 250-F113 Revision 7 Page 2 of 2





NAME OF ACCUSED

COMMONWEALTH OF VIRGINIA DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF COURT

BREATH ANALYSIS SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M DFS LICENSE NUMBER 18910 TEST EQUIPMENT NUMBER 010453	AGENCY DFS Central Lab DATE TEST CONDUCTED 02/20/2014
STANTON, HEATHER, M DFS LICENSE NUMBER LICENSE EXPIRES 18910 TEST EQUIPMENT NUMBER	DFS Central Lab DATE TEST CONDUCTED
DFS LICENSE NUMBER 18910 TEST EQUIPMENT NUMBER LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED
18910 10/01/2014 TEST EQUIPMENT NUMBER	
TEST EQUIPMENT NUMBER	02/20/2014
010453	
V V. V. V.	
RESULTS: TIME SAMPLE TAKEN11:47 EST	
SAMPLE'S ALCOHOL CONTENT 0.00	GRAMS PER 210 LITERS OF BREATH
ATTEST:	
EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROV CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE	ST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF ZED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS IT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND
BREATH TEST OPERATOR	
\square I HAVE RECEIVED A COPY OF THIS CERTIFICATE O	F ANALYSISSUBJECT'S SIGNATURE
\square SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFIC	CATE OF ANALYSISOPERATOR'S SIGNATURE



IntoxNet MIS Report Report Generated 24 Feb 2014 at 08:56

Test Results

Instrument Serial Number 010453

Subject Test Test # 000294 Test Location 1 Department of

Test Date 20 Feb 2014

Operator's Last Name STANTON

Agency DFS Central Lab Card Serial Number 118910

Subject's Last Name INSTRUMENT

Subject's Middle Initial

Driver's License Number

Driver's License State

Test Location 2 Forensic Science

Subject's Date of Birth 00/00/0000

Test Time 11:41

Remote/Local Local

Operator's First Name HEATHER

System Check Passed

Test Location 3

Operator's Middle Initial M

License Number 18910

Effective Date 10/01/2012 Expiration Date 10/01/2014 Subject's First Name TEST

Subject's Sex Male

Driver's License Expiration 00/00/0000

Court Name DFS

End Date 20 Feb 2014 End Time 11:48 Result Time 11:47 Result Date 20 Feb 2014 Result 0.00 Data Type DIAG Sample Value Pass Sample Time 11:41 Data Type BLK Sample Value 0.000 Sample Time 11:42 Data Type CHK Sample Value 0.099 Sample Time 11:42 Data Type BLK Sample Value 0.000 Sample Time 11:44 Data Type SUBJ Sample Value 0.000 Sample Time 11:44 Data Type BLK Sample Value 0.000 Sample Time 11:45 Data Type SUBJ Sample Value 0.000 Sample Time 11:47

Sample Value 0.000

Data Type BLK Standard Type Dry Gas Std Standard Lot Number AG219101-11

Tank Pressure 154

Blow Sample Number 1 Blow Sample Number 2 Blow Duration 3.39 sec Blow Duration 3.37 sec

Barometric Pressure 735 mmHg Blow Volume 1848 cc Blow Volume 1751 cc

End-of-Blow Time 11:44 End-of-Blow Time 11:47

Sample Time 11:48

Test Status Code 0

Standard Value 0.097

Standard Expiration Date 07/09/2014

Tamper Evident Stamp 88f7c315

Test Status Success

8rd

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010453 Test Number: 295
Test Date: 02/20/2014 Test Time: 11:49 EST

Dry Gas Target: 0.097

Lot Number: AG219101-11 Exp Date: 07/09/2014

Tank Pressure: 154 psi Barometric Pressure: 735 mmHg

System Check: Passed

Test	g/210L	Time
BLK	ō.000	11:50
CHK	0.099	11:50
BLK	0.000	11:52
CHK	0.099	11:53
BLK	0.000	11:54
CHK	0.099	11:55

Test Status: Success

Calibration CRC: 60D5B217

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010453 Test Number: 296

Dry Gas Target: 0.290

Lot Number: AG322702-05 Exp Date: 08/15/2015

Tank Pressure: 458 psi Barometric Pressure: 735 mmHg
System Check: Passed

Test	g/210L	Time
BLK	0.000	11:58
CHK	0.296	11:59
BLK	0.000	12:01
CHK	0.296	12:02
BLK	0.000	12:04
СНК	0.296	12:05

Test Status: Success

Calibration CRC: 60D5B217



CLIFTON FORGE POLICE DEPARTMENT

B. G. BALSER, CHIEF OF POLICE

547 MAIN STREET CLIFTON FORGE, VIRGINIA 24422 (540) 863-2513 • FAX (540) 863-2537

January 13, 2014

010453 Hms H2814

Department of Forensic Science Breath-Alcohol Program 700 North Sth Street Richmond, Virginia 23219

Dear Sir/Madam,

As you are aware, the Clifton Forge Police Department currently utilizes an Intox EC/IR II, s/n 010453, which is housed in within our booking and processing area. We are currently in the process of making some changes to the area and would like to relocate the Intoximeter to another room which is located in an adjacent area. We have already had the room equipped with a dedicated computer line designated for the unit.

Moving of the unit will require not only disassembly of the intoximeter, but disassembly of the "cage" that it is housed in as well. This is a simple operation and we will have someone on hand to move the "cage" and to assist your technician in the moving of your unit.

I therefore ask if you could make arraignments to have the intoximeter unit relocated at a time that is suitable to your department.

I thank you for your attention to this request and please feel free to contact me at 540 863-2513 to discuss this matter as needed.

Sincerely,

Barry G Balser

Chief of Police